



## Quality Control Certificate

**Product name:** LilliTest MG positive serum (freeze dried)

**Product Description:** *Mycoplasma gallisepticum* positive serum for the use as positive control in MG RPA Test

**Product Presentation:** 1.5 ml (freeze dried)

**Product Code:** V2026

**Batch Number:** MGPOS02161

**Production Date:** February 2016

**Expiry Date:** February 2018

**Storage:** +2 and +8°C

Does not contain any preservatives.

### Preparation of the serum before use

- Reconstitute with **1.5ml of sterile NaCl for 9g/l** in aseptic conditions.
- Leave the product to slowly rehydrate for at least 2 minutes.
- After rehydration keep the serum between +2 and +8°C for three months.

**Quality Control**  
**Lillidale Diagnostics**

Date: 26 February 2016