



Quality Control Certificate

Product name: LilliTest MM positive serum (freeze dried)

Product Description: *Mycoplasma meleagridis* positive serum for the use as positive control in MM RPA Test

Product Presentation: 2.5 ml (freeze dried)

Product Code: V2028

Batch Number: MMPOS9151

Production Date: September 2015

Expiry Date: September 2016

Storage: +2 and +8°C

Does not contain any preservatives.

Preparation of the serum before use

- Reconstitute with **2.5ml of sterile NaCl for 9g/l** in aseptic conditions.
- Leave the product to slowly rehydrate for at least 2 minutes.
- Centrifuge the rehydrated product for 15 minutes at 4000g and remove the supernatant immediately.
- After rehydration keep the serum between +2 and +8°C for three months.

Quality Control
Lillidale Diagnostics

Date: 4 September 2015